



AB GERMA

EC Declaration of conformity no.: 071004-002 (issue 2)

Product Name:	Set of Splints, type FERNO Washington UK
Ferno Article No:	AS 190
Germa Article No:	23005002210
Description:	Vacuum splints are tools for fixation and immobilisation of patients with suspected injuries on arm, leg, shoulder or hip. They can also be used as vacuum mattresses for children. The splint consists of an airtight shell with fixation straps, vacuum valve and polystyrene granules. Set of Splints comprising of: 1 x 23005002012, AS 100, Arm splint, short 1 x 23005002013, AS 120, Arm splint 1 x 23005002014, AS 140, Leg splint 1 x 23005002026, AS 180, Bag for splints 1 x 23005002017 Repair kit FW 1 x 23272000711, Vacuum Hand Pump
Accessory:	

Identification: All products with serial numbers issued after LOT number: **030704**,
date: **2003-07-04** (yyyy-mm-dd).

We the manufacturer hereby declare that this product is in conformity with the requirements in:

- Annex I and Annex VII and Annex IX in Council directive 93/42/EEC.
Equipment class: I, non-sterile, no measuring function.
- Swedish Medical Devices Act 1993:584

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position: Quality Manager
Name: Johan Attelid

Date: 19/12-2011

Sign:

Postadress	Adress	Telefon	Telefax	Organisationsnummer. F-skattebevis	Bankgiro	Postgiro
Box 2127 S-291 02 KRISTIANSTAD SWEDEN e-mail: info@germa.se	Industrigatan 54-56 KRISTIANSTAD	+46 (0)44-123030	+46 (0)44-103179	556086-9413	481-4125	21 68 22-7